

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

JUL 22 2010

| | | | |
|---|--|---------------------------------------|---------------------------------------|
| Supervisor/Operator Performing the Verification Procedure: | | DataMaster cdm S/N 130119 ✓ | |
| Name <u>Tobin P Brennan</u> | | ID# <u>4542</u> ✓ | Date <u>7-20-10</u> ✓ |
| A | Agency <u>Soldotna Police Department</u> | | Phone # <u>(907) 262-4455</u> |
| Instrument Location <u>Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669)</u> | | | |
| | | | |
| B | Alco S/N <u>X172985</u> ✓ | Target Value <u>.088</u> ✓ | High Pressure <u>800</u> |
| Alco Test Values | | <u>.088</u> ✓ 1 st Alco | <u>.088</u> ✓ 2 nd Alco |
| Signature <u>[Signature]</u> | | [Signature] 7/23/10 | |
| (OVER) | | | |

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 9th day of AUGUST, 2010.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

JULY 20, 2010

OPERATOR'S NAME:

BRENNAN

OPERATOR'S NUMBER: 4542

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/NI :

VOC

O.L. #: VOC

DEPT/AGENCY: SLD1

CASE/REPORT: VOC

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: X172985

— BREATH ANALYSIS —

.000 ADJUSTED FOR 30.01 in

ALCO TARGET .000

BLANK TEST .000

INTERNAL STANDARD VERIFIED

ALCO TV 29.99 in .000

BLANK TEST .000

SUBJECT SAMPLE .000

BLANK TEST .000

ALCO TV 30.01 in .000

BLANK TEST .000

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

JULY 20, 2010

TIME 22:09

— DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 03/29/00

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 45c

BAROMETER: 29.99 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢ £ ¤ ¥ ¦ § ¨ © ª « ¬ ® ¯ ° ± ² ³ ´ µ ¶ · ¸ ¹ º » ¼ ½ ¾